

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	ASH WASTE
SITE REFERENCE:	
SITE ADDRESS:	MILLS RD BRETTON CHESHIRE
POST CODE:	CH4 0BY
WEIGHBRIDGE LOCATION:	MAIN WEIGHBRIDGE
WEIGHBRIDGE SIZE:	15 X 3

SAP REPORT NO.:	4442134
INDICATOR TYPE:	I 200
INDICATOR SERIAL NO.:	28F630403
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	DRY

REPEATABILITY TEST

Direction →
of
Travel ←

43,530 kg
Weighbridge Viewed from Indicator Position
Indicator Reading: 43,530 kg

Error Allowed
+/-2dvs or +/-40kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg) 10,000

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
3,900		10,000	10,000	10,000	0
	ERROR (e)	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Allowed Error
0	0	0	0.0	Pass	
5,000	5,000	5,000	0.0	Pass	error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass	error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass	error +/-1dvs or 20kg
Substitute Load	19,400	19,400	0.0	Pass	error +/-2dvs or 40kg
25,000	24,520	24,520	0.0	Pass	error +/-2dvs or 40kg
35,000	34,520	34,520	0.0	Pass	error +/-2dvs or 40kg
40,000	39,520	39,520	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	39,920	39,920	0.0	Pass	error +/-2dvs or 40kg
44,000	43,530	43,530	0.0	Pass	error +/-3dvs or 60kg

MAINT LOG UPDATED: YES

MAXIMUM RANGE TESTED: NO

TEST EQUIPMENT USED: Test Weights Traceable to National Standards

CALIBRATION DATE: WEDNESDAY 18 JULY 2018.

AUTHORISED ENGINEER:

CERTIFICATE NUMBER:

NEXT CALIBRATION DATE:

D DUXBURY

M8882

18 JULY 2019.

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name		Date	18/07/2018
Signature		Engineer's Signature	

We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.

v14. May 2016