

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	ASH WASTE
SITE REFERENCE:	
SITE ADDRESS:	REDWITHER RD WREXHAM CHESHIRE
POST CODE:	LL13 9RD
WEIGHBRIDGE LOCATION:	MAIN WEIGHBRIDGE
WEIGHBRIDGE SIZE:	15 X 3

SAP REPORT NO.:	4442015
INDICATOR TYPE:	EV 2002
INDICATOR SERIAL NO.:	85604
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	DRY

REPEATABILITY TEST

Direction of Travel →
←

43,540 kg
Weighbridge Viewed from Indicator Position
Indicator Reading: 43,540 kg

Error Allowed
+/-2dvs or +/-40kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	10,000
-------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
3,900		10,000	10,000	10,000	0
	ERROR (e)	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Allowed Error
0	0	0	0.0	Pass	error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass	error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass	error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	19,400	19,400	0.0	Pass	error +/-2dvs or 40kg
25,000	24,340	24,340	0.0	Pass	Added weight (kg) 5,000 error +/-2dvs or 40kg
35,000	34,340	34,340	0.0	Pass	10,000 error +/-2dvs or 40kg
40,000	39,340	39,340	0.0	Pass	5,000 error +/-2dvs or 40kg
Substitute Load	39,740	39,740	0.0	Pass	0 error +/-2dvs or 40kg
44,000	43,540	43,540	0.0	Pass	5000 error +/-3dvs or 60kg

MAINT LOG UPDATED: YES

MAXIMUM RANGE TESTED: NO

TEST EQUIPMENT USED: Test Weights Traceable to National Standards
CALIBRATION DATE: WEDNESDAY 18 JULY 2018.

AUTHORISED ENGINEER: D DUXBURY
CERTIFICATE NUMBER: M8882
NEXT CALIBRATION DATE: 18 JULY 2019.

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name	M. K. K.	Date	18/07/2018
Signature	<i>[Signature]</i>	Engineer's Signature	<i>[Signature]</i>

We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.