

**E.U. WEIGHBRIDGE VERIFICATION CERTIFICATE – No: 5502**

<b>CUSTOMER:</b> Ash Waste	<b>TYPE APPROVAL CERTIFICATE No:</b> LNE-14588 Rev.3
<b>SITE ADDRESS:</b> Broughton Mills Road	<b>INDICATOR TYPE:</b> X201-B (i200)
Bretton	<b>INDICATOR TEST CERTIFICATE No:</b> LNE 01-03 Rev. 2
Flintshire	<b>INDICATOR SERIAL No:</b> 28F630403
CH4 0BY	<b>MODULE TEST CERTIFICATE No:</b> LNE TC 05-02 Rev 1
<b>MANUFACTURER:</b> Precia Molen	<b>MAXIMUM CAPACITY:</b> 50,000 kg
<b>WEIGHBRIDGE TYPE:</b> VS400	<b>DIVISION (e):</b> 20 kg
<b>PLATFORM SIZE:</b> 15m x 3m	<b>LOADCELL TEST CERTIFICATE No:</b> PTB D09-06.08
<b>EST. PLATFORM WEIGHT:</b> 9,500 kg	<b>PRINTER SERIAL No:</b> N/A
<b>WEIGHBRIDGE LOCATION:</b> Main Yard Weighbridge	<b>TARE FACILITY:</b> Disabled
<b>INITIAL VERIFICATION:</b> <input type="checkbox"/>	<b>RE-VERIFICATION:</b> <input checked="" type="checkbox"/>

**Repeatability Test - MPE 0.3e**

	Indication (HR)	Indication (HR)	Indication (HR)
Loaded WTU	43784	43784	43786
	PASS	✓	FAILED

**Accuracy of Zero, Linearity/Hysteresis, & Comparison Tests**

Approximate Test Interval	MPE	Actual Load	Indication Up (HR)	Error	Corrective Error	Indication Down (HR)	Error	Corrective Error	Comparison
Zero	0.25	0	2	0.1e	0.1e	2	0.1e	0e	
2e	0.5	40	42	0.1e	0e	44	0.2e	0.1e	
20e	0.5	440	442	0.1e	0e	444	0.2e	0.1e	
500e	1.0	10040	10044	0.2e	0.1e	10044	0.2e	0.1e	
1000e	1.0	20040	20044	0.2e	0.1e				
Substitute Load	1.0	19392	19396	0.2e	0.1e	19394	0.1e	0e	Yes
1250e	1.0	24392	24396	0.2e	0.1e	24396	0.2e	0.1e	
1500e	1.0	29392	29394	0.1e	0e	29396	0.2e	0.1e	
2000e	1.0	39392	39398	0.3e	0.2e	39396	0.2e	0.1e	
Substitute Load	1.0	39342	39348	0.3e	0.2e	39346	0.2e	0.1e	
2200e	1.5	43742	43748	0.3e	0.2e				
PASSED ✓					FAILED				

**Loadcell Positions**

2	3	4
1	6	5

DWI

FLT Substitute Load  
**3,948 kg**


**Comparison Test**

Printer	
Remote Display	
PC	✓
Other	
Pass	✓
Fail	

**Eccentric Load Test – Up to 500e MPE = 0.5e, over 500e MPE = 1.0e**

Position	1	2	3	4	5	6	7	8	9	10
Test Load	10040	10040	10040	10040	10040	10040				
Indication (HR)	10044	10042	10044	10044	10044	10044				
Error	0.2e	0.1e	0.2e	0.2e	0.2e	0.2e				
PASSED ✓					FAILED					

<b>ENVIRONMENTAL CONDITIONS:</b> Dry	<b>AUTHORISED ENGINEER:</b> C.Swinburne
<b>TEST WEIGHTS USED:</b> Precia-Molen UK Ltd	<b>CERTIFICATE NO:</b> W18034 & 70591903
<b>VERIFICATION DATE:</b> 22-10-18	<b>NEXT AFC DATE :</b> 22-10-19

Signature:  Date: .....29/10/2018.....

The EU Weighbridge Verification Certificate complies with the requirements of NAWI directive 2014/31/EU and EN45501:2015. The tests are also in compliance with the UK Weighing Federation Code of Practice for Non Automatic Weighing Instruments.  
EU Weighbridge Verification Certificate. – Issue: 16.1 Dated: 03/07/2018