

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	Ash Waste
SITE REFERENCE:	
SITE ADDRESS:	Station Rd. Rowley Regis
POST CODE:	B65 0JU
WEIGHBRIDGE LOCATION:	Main Weighbridge
WEIGHBRIDGE SIZE:	16 x 3

REPORT NO.:	4610449
INDICATOR TYPE:	System 2X
INDICATOR SERIAL NO.:	145400
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	10
Weather Conditions	Very Wet

REPEATABILITY TEST

Direction of Travel →
←

Indicator Reading:	43,900 kg
Weighbridge Viewed from Indicator Position	
Indicator Reading:	43,900 kg

Error Allowed
+/-3dvs or +/-60kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	19,730
-------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
3,760		19,720	19,720	19,720	0
	ERROR (e)	error +/-2dvs or 40kg		error +/-2dvs or 40kg error +/-1dvs or 20kg	
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass		error +/-2dvs or 40kg
24,400	24,160	24,160	0.0	Pass		error +/-3dvs or 60kg
Substitute Load	23,490	23,490	0.0	Pass		error +/-3dvs or 60kg
28,600	28,490	28,490	0.0	Pass	5,000	error +/-3dvs or 60kg
38,600	38,490	38,500	1.0	Pass	10,000	error +/-3dvs or 60kg
43,600	43,490	43,500	1.0	Pass	5,000	error +/-3dvs or 60kg
44,000	43,890	43,900	1.0	Pass	400	error +/-3dvs or 60kg

MAINT LOG UPDATED:

MAXIMUM RANGE TESTED:

TEST EQUIPMENT USED: Test Weights Traceable to National Standard

CALIBRATION DATE: 25 August 2020

AUTHORISED ENGINEER:
CERTIFICATE NUMBER:
NEXT CALIBRATION DATE:

Kevin Delve
TR10874/TR10875
24 August 2021

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name	M. KIRK	Date	25/08/2020
Signature		Engineer's Signature	

We Certify that the Weighbridge Calibration check complies to the current Weights & Measures criteria and the calibration procedure is carried out in accordance with the Weighing Federation's Code of practice.

(V13LH)