

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	Ash Waste
SITE REFERENCE:	
SITE ADDRESS:	Redwither Rd. Wrexham Industrial Estate Wrexham
POST CODE:	LL13 9RD
WEIGHBRIDGE LOCATION:	Main Weighbridge
WEIGHBRIDGE SIZE:	16 x 3

REPORT NO.:	4510717
INDICATOR TYPE:	EV2002
INDICATOR SERIAL NO.:	85604
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	Sunny

REPEATABILITY TEST

Direction →
of
Travel ←

Indicator Reading:	43,740 kg
Weighbridge Viewed from Indicator Position	
Indicator Reading:	43,740 kg

Error Allowed
+/-3dvs or +/-60kg
Pass

END - MIDDLE - END - ZERO TEST

Fork Lift Truck used as sub load.	APPLIED LOAD (kg)	19,640
-----------------------------------	-------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
3,740		19,620	19,620	19,640	0
	ERROR (e)	error +/-2dvs or 40kg		error +/-2dvs or 40kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass		error +/-1dvs or 20kg
24,400	24,140	24,120	-1.0	Pass		error +/-2dvs or 40kg
Substitute Load	23,380	23,360	-1.0	Pass		error +/-2dvs or 40kg
28,600	28,380	28,360	-1.0	Pass	5,000	error +/-2dvs or 40kg
38,600	38,380	38,340	-2.0	Pass	10,000	error +/-2dvs or 40kg
43,600	43,380	43,340	-2.0	Pass	5,000	error +/-3dvs or 60kg
44,000	43,780	43,740	-2.0	Pass	400	error +/-3dvs or 60kg

MAINT LOG UPDATED: No

MAXIMUM RANGE TESTED: No

TEST EQUIPMENT USED: Test Weights Traceable to National Standard

CALIBRATION DATE: 24 April 2020

AUTHORISED ENGINEER:

Kevin Delve

CERTIFICATE NUMBER:

TR10874/TR10875

NEXT CALIBRATION DATE:

23 April 2021

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name	Daniel Flood	Date	24/04/2020
Signature		Engineer's Signature	

We Certify that the Weighbridge Calibration check complies to the current Weights & Measures criteria and the calibration procedure is carried out in accordance with the Weighing Federation's Code of practice.

(v13LH)

