

**VERIFICATION CERTIFICATE – No: 6861**

**CUSTOMER:** ALANS SKIP HIRE - ASH  
**SITE ADDRESS:** Broughton Industrial Estate  
 Broughton Mills Road  
 Bretton, Chester  
 CH4 0BY

**MANUFACTURER:** Weightron Bilanciai  
**WEIGHBRIDGE TYPE:** Steel Surface Mount  
**PLATFORM SIZE:** 18m x 3m  
**EST. PLATFORM WEIGHT:** 12,000 kg  
**WEIGHBRIDGE LOCATION:** Main Weighbridge

**CUSTOMER EMAIL ADDRESS:** melaniemorris@ashgrouppltd.co.uk  
**INITIAL VERIFICATION:**  **RE-VERIFICATION:**

**TYPE APPROVAL CERTIFICATE No:**  
**INDICATOR TYPE:**  
**INDICATOR TEST CERTIFICATE No:**  
**INDICATOR SERIAL No:**  
**MODULE TEST CERTIFICATE No:**  
**MAXIMUM CAPACITY:**  
**DIVISION (e):**  
**LOADCELL TEST CERTIFICATE No:**  
**PRINTER SERIAL No:**  
**TARE FACILITY:**

LNE-14588 Rev.4  
 X201-A (X201-B i200)  
 LNE 01-03 Rev. 2  
 28F630403  
 LNE 05-02 Rev 1  
 50,000 kg  
 20 kg  
 PTB D09-06.08  
 N/A  
 Disabled

**Repeatability Test - MPE 0.3e**

	Indication (HR)	Indication (HR)	Indication (HR)
Loaded WTU	43420	43420	43420
	PASS	✓	FAILED

**Accuracy of Zero, Linearity/Hysteresis, & Comparison Tests**

Approximate Test Interval	MPE	Actual Load	Indication Up (HR)	Error	Corrective Error	Indication Down (HR)	Error	Corrective Error	Comparison
Zero	0.25	0	2	0.1e	0.1e	0	0e	-0.1e	
2e	0.5	40	42	0.1e	0e	40	0e	-0.1e	
20e	0.5	440	444	0.2e	0.1e	442	0.1e	0e	
500e	1.0	10040	10044	0.2e	0.1e	10044	0.2e	0.1e	
1000e	1.0	20040	20044	0.2e	0.1e				
Substitute Load	1.0	19092	19096	0.2e	0.1e	19096	0.2e	0.1e	Yes
1250e	1.0	24092	24094	0.1e	0e	24094	0.1e	0e	
1500e	1.0	29092	29088	-0.2e	-0.3e	29088	-0.2e	-0.3e	
2000e	1.0	39092	39084	-0.4e	-0.5e	39084	-0.4e	-0.5e	
Substitute Load	1.0	38440	38432	-0.4e	-0.5e	38432	-0.4e	-0.5e	
2200e	1.5	43440	43430	-0.5e	-0.6e				
PASSED ✓					FAILED				

**Loadcell Positions**

2	3	4
1	6	5
DWI		

FLT Substitute Load  
**4,348 kg**

**Comparison Test**

Printer	
Remote Display	
PC	✓
Other	
Pass	✓
Fail	

**Eccentric Load Test – Up to 500e MPE = 0.5e, over 500e MPE = 1.0e**

Position	1	2	3	4	5	6	7	8	9	10
Test Load	10040	10040	10040	10040	10040	10040				
Indication (HR)	10044	10036	10038	10040	10036	10040				
Error	0.2e	-0.2e	-0.1e	0e	-0.2e	0e				
Corrective Error	0.1e	-0.3e	-0.2e	-0.1e	-0.3e	-0.1e				
PASSED ✓					FAILED					

**ENVIRONMENTAL CONDITIONS:** Icy cold, dry  
**TEST WEIGHTS USED:** Precia-Molen UK Ltd  
**VERIFICATION DATE:** 20/01/2023

**AUTHORISED ENGINEER:** Darren Entwistle  
**CERTIFICATE NO:** TR 11761 & TR 11762/TR 11714  
**NEXT CHECK DATE:** 20/01/2024

Signature: ..... *D. Entwistle* .....

Date: ...27/01/2023.....

*This Verification Certificate complies with the requirements of Non-Automatic Weighing Instrument Regulations 2016 No: 1152 and EN45501:2015. The tests are also carried out in compliance with the UK Weighing Federation Code of Practice for Non Automatic Weighing Instruments. Form 080 UK Verification V1.0*