

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	REPORT NO.: 4902360
SITE ADDRESS:	INDICATOR TYPE: System X2
POST CODE:	INDICATOR SERIAL NO.: 14500
WEIGHBRIDGE LOCATION: Main Weighbridge	MAXIMUM CAPACITY (kg): 50,000
WEIGHBRIDGE SIZE: 16x3	DIVISION 'e' (kg): 20
	Weather Conditions: wet

REPEATABILITY TEST

Direction →
of
Travel ←

Indicator Reading:	43,520 kg
Weighbridge Viewed from Indicator Position	
Indicator Reading:	43,520 kg

Error Allowed
+/-2dvs or +/-40kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	10,000
--------------------------	--------

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
		4,260	10,020	10,000	10,000
	ERROR (e)	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Allowed Error
0	0	0	0.0	Pass	error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass	error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass	error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	19,340	19,340	0.0	Pass	error +/-2dvs or 40kg
25,000	24,340	24,340	0.0	Pass	error +/-2dvs or 40kg
35,000	34,340	34,340	0.0	Pass	error +/-2dvs or 40kg
40,000	39,340	39,340	0.0	Pass	error +/-2dvs or 40kg
Substitute Load	38,520	38,520	0.0	Pass	error +/-2dvs or 40kg
44,000	43,520	43,520	0.0	Pass	error +/-3dvs or 60kg

MAINT LOG UPDATED:

MAXIMUM RANGE TESTED:

TEST EQUIPMENT USED: Test Weights Traceable to National Standards
CALIBRATION DATE: 12 December 2023

AUTHORISED ENGINEER: Allan Macintyre
TEST WEIGHT CERTIFICATE NUMBER: TR11952TR11953
NEXT CALIBRATION DATE: 10 December 2024

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Name	<i>Dave Rocher</i>	Date	12 12. 2023
Customer Signature	<i>[Signature]</i>	Engineer's Signature	<i>[Signature]</i>

We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.