

## WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

<b>CUSTOMER:</b>	Ash Waste Services Ltd
<b>SITE ADDRESS:</b>	Redwither Rd
	Wrexham Ins Est
	Wrexham
<b>POST CODE:</b>	LL13 9RD
<b>WEIGHBRIDGE LOCATION:</b>	Main W/B
<b>WEIGHBRIDGE SIZE:</b>	15 X 3

<b>REPORT NO.:</b>	40018542
<b>INDICATOR TYPE:</b>	Dini Argeo
<b>INDICATOR SERIAL NO.:</b>	4514815
<b>MAXIMUM CAPACITY (kg):</b>	50,000
<b>DIVISION 'e' (kg):</b>	20
<b>Weather Conditions</b>	Overcast And Cool

### REPEATABILITY TEST

Direction →  
of  
Travel ←

<b>Indicator Reading:</b>	43,980 kg
Weighbridge Viewed from Indicator Position	
<b>Indicator Reading:</b>	43,980 kg

Error Allowed  
+/-2dvs or +/-40kg  
**Pass**

### END - MIDDLE - END - ZERO TEST

<b>APPLIED LOAD (kg)</b>	10,000
--------------------------	--------

		End	Middle	End	Zero
<b>Fork Lift Truck Weight (kg)</b>	<b>INDICATOR READING (kg)</b>	9,980	9,980	10,000	0
<b>4,660</b>	<b>ERROR (e)</b>	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg	error +/-1dvs or 20kg
	<b>PASS/FAIL</b>	Pass	Pass	Pass	Pass

### RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Allowed Error
0	0	0	0.0	Pass	error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass	error +/-1dvs or 20kg
10,000	10,000	9,980	-1.0	Pass	error +/-1dvs or 20kg
20,000	20,000	19,980	-1.0	Pass	error +/-2dvs or 40kg
Substitute Load	19,340	19,320	-1.0	Pass	error +/-2dvs or 40kg
25,000	24,340	24,320	-1.0	Pass	error +/-2dvs or 40kg
35,000	34,340	34,300	-2.0	Pass	error +/-2dvs or 40kg
40,000	39,340	39,300	-2.0	Pass	error +/-2dvs or 40kg
Substitute Load	39,000	38,960	-2.0	Pass	error +/-2dvs or 40kg
44,000	44,000	43,980	-1.0	Pass	error +/-3dvs or 60kg

<b>MAINT LOG UPDATED:</b>	NO
<b>MAXIMUM RANGE TESTED:</b>	NO
<b>TEST EQUIPMENT USED:</b>	Test Weights Traceable to National Standards
<b>CALIBRATION DATE:</b>	17 September 2024

<b>AUTHORISED ENGINEER:</b>	Simon Woodman
<b>TEST WEIGHT CERTIFICATE NUMBER:</b>	TR12234/TR12235
<b>NEXT CALIBRATION DATE:</b>	16 September 2025

<b>Comments</b>

*We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction*

Name		Date	17/09/2024
Customer Signature		Engineer's Signature	

*We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.*