



WEIGHBRIDGE SERVICES

The Chapel, Carrwell Lane, Sheffield, S6 1NR, United Kingdom
Tel: +44 (0)114 321 6364 / Email: info@weighbridgeservices.com

CERTIFICATE OF CALIBRATION

| | | | |
|------------------------|-------|--------------|------------|
| Certificate No: | J5829 | Date: | 26/09/2025 |
|------------------------|-------|--------------|------------|

| | | | |
|------------------------|----------------------|--|--|
| Company Name: | Ash Waste Ltd | | |
| Address: | Broughton Mills Road | | |
| | Bretton | | |
| | Flintshire | | |
| | | | |
| | CH4 0BY | | |
| Contact Name: | | | |
| Contact Number: | | | |

| | | | |
|----------------------|-------------------------|-----------------------|--------------|
| ID Number: | <i>Main Weighbridge</i> | | |
| Machine Type: | Weighbridge | Manufacturer: | Precia Molen |
| Model Number: | I 30 | Serial Number: | 06F1116924 |
| Capacity: | 50,000 KG | Increments: | 20KG |

| | |
|---|---------|
| Owner of Class M1 Test Weights Used: | D Brash |
| Certificate Number: | W25012 |
| Certificate Number: | W25012 |

Number of load points (please tick):

| | | | |
|-----------------------|-----|--------|-----|
| <input type="radio"/> | 2 | | 3 |
| | END | MIDDLE | END |
| | 1 | | 4 |

| | | | | | |
|----------------------------------|-----|--|--------|--|-----|
| <input checked="" type="radio"/> | 2 | | 6 | | 3 |
| | END | | MIDDLE | | END |
| | 1 | | 5 | | 4 |

| | | | | | | | |
|-----------------------|-----|--|--------|--|-----|--|---|
| <input type="radio"/> | 2 | | 6 | | 7 | | 3 |
| | END | | MIDDLE | | END | | |
| | 1 | | 5 | | 8 | | 4 |

All Calibrations are performed to UKWF Guidelines

| Eccentricity/Corners | | | | | |
|----------------------|---------------------|---------------|------------|-----------------|------------|
| Load Position | Weight Applied (kg) | As Found (kg) | Error (kg) | Definitive (kg) | Error (kg) |
| 1 | 6,000kg | 6,000kg | 0 kg | 6,000kg | 0 kg |
| 2 | 6,000kg | 6,000kg | 0 kg | 6,000kg | 0 kg |
| 3 | 6,000kg | 6,000kg | 0 kg | 6,000kg | 0 kg |
| 4 | 6,000 kg | 6,000kg | 0 kg | 6,000kg | 0 kg |
| 5 | 6,000kg | 6,000kg | 0 kg | 6,000kg | 0 kg |
| 6 | | | | | |
| 7 | | | | | |
| 8 | | | | | |

| Load Position | As Found (kg) | Definitive (kg) | Make-up Weights Used: |
|---------------|---------------|-----------------|---|
| END | 10,000 | 10,000 | Substitute 1 - Wagon kg – 18,380 kg Substitute 2 -Forklift kg – 5,940 kg |
| MIDDLE | 10,000 | 10,000 | |
| END | 10,000 | 10,000 | |

| Linearity | | | | | |
|-------------------------------------|---------------|------------|-----------------|------------|------------|
| Load Applied (kg) | As Found (kg) | Error (kg) | Definitive (kg) | Error (kg) | Error (kg) |
| 0 kg | 0 kg | 0 kg | 0 kg | 0 kg | 0 kg |
| 5,000kg | 5,000kg | 0 kg | 2,000kg | 0 kg | 0 kg |
| 10,000kg | 10,000 kg | 0 kg | 10,000kg | 0 kg | 0 kg |
| 20,000kg | 20,020 kg | +20 kg | 20,020kg | +20 kg | +20 kg |
| Sub 1 – (Wagon) | 18,380 kg | +20 kg | 18,380 kg | +20 kg | +20 kg |
| Sub1 + 10,000kg | 28,380 kg | +20kg | 28,380g | +20kg | +20kg |
| Sub 1 + 20,000kg | 38,360 kg | 0 kg | 38,360kg | 0 kg | 0 kg |
| Sub 1 + 20,000kg + Sub 2 (Forklift) | 44,320 kg | +20 kg | 44,320 kg | +20 kg | +20 kg |

| Repeatability | | | Weight and Measures tolerances for use with a Class III Machine: 0 - 499 Divisions = +/- 1 Division 500 - 1999 Divisions = +/- 2 Divisions 2000 - 3000 Divisions = +/- 3 Divisions |
|---------------|---------------|-----------------|---|
| Applied Load | As Found (kg) | Definitive (kg) | |
| 20,000 kg | 20,020 kg | 20,020 kg | |
| 20,000 kg | 20,020kg | 20,020 kg | |
| 20,000 kg | 20,020kg | 20,020 kg | |

| | |
|----------------------------------|------------|
| Engineer: | C. Koning |
| Calibration Date: | 26/09/2025 |
| Calibration Renewal Date: | Sept 2026 |



All Calibrations are performed to UKWF Guidelines