



WEIGHBRIDGE SERVICES

The Chapel, Carrwell Lane, Sheffield, S6 1NR, United Kingdom
Tel: +44 (0)114 321 6364 / Email: info@weighbridgeservices.com

CERTIFICATE OF CALIBRATION

Certificate No:	J5830	Date:	26/09/2025
------------------------	-------	--------------	------------

Company Name:	Ash Waste Ltd		
Address:	Redwither Road		
	Wrexham		
	LL13 9RD		
Contact Name:			
Contact Number:			

ID Number:	<i>Main Weighbridge</i>		
Machine Type:	Weighbridge	Manufacturer:	Dini Argeo
Model Number:	3590 EGT	Serial Number:	04514815
Capacity:	50,000 KG	Increments:	20KG

Owner of Class M1 Test Weights Used:	D Brash
Certificate Number:	W25012
Certificate Number:	W25012

Number of load points (please tick):

<input checked="" type="checkbox"/>	2			3
	END	MIDDLE	END	
	1			4

<input type="checkbox"/>	2		6		3
	END	MIDDLE	END		
	1		5		4

<input type="checkbox"/>	2		6		7		3
	END	MIDDLE	END				
	1		5		8		4

All Calibrations are performed to UKWF Guidelines

Eccentricity/Corners					
Load Position	Weight Applied (kg)	As Found (kg)	Error (kg)	Definitive (kg)	Error (kg)
1	6,000kg	6,000kg	0 kg	6,000kg	0 kg
2	6,000kg	6,000kg	0 kg	6,000kg	0 kg
3	6,000kg	6,000kg	0kg	6,000kg	0 kg
4	6,000 kg	6,000kg	0 kg	6,020 kg	0 Kg
5					
6					
7					
8					

Load Position	As Found (kg)	Definitive (kg)	Make-up Weights Used:
END	10,000	10,000	Substitute 1 - Wagon kg – N/A Substitute 2 -Forklift kg – 5,980 kg
MIDDLE	10,000	10,000	
END	10,000	10,000	

Linearity					
Load Applied (kg)	As Found (kg)	Error (kg)	Definitive (kg)	Error (kg)	Error (kg)
0 kg	0 kg	0 kg	0 kg	0 kg	0 kg
5,000kg	5,000kg	0 kg	2,000kg	0 kg	0 kg
10,000kg	10,020 kg	+20 kg	10,020kg	+20 kg	+20 kg
20,000kg	20,000 kg	0 kg	20,000kg	0 kg	0 kg
Sub 1 – (Wagon)	18,380 kg	0 kg	18,780 kg	0 kg	0 kg
Sub1 + 10,000kg	28,380 kg	0kg	28,380kg	0kg	0kg
Sub 1 + 20,000kg	38,380 kg	0 kg	38,380kg	0 kg	0 kg
Sub 1 + 20,000kg + Sub 2 (Forklift)	44,360 kg	0 kg	44,360 kg	0 kg	0 kg

Repeatability			Weight and Measures tolerances for use with a Class III Machine: 0 - 499 Divisions = +/- 1 Division 500 - 1999 Divisions = +/- 2 Divisions 2000 - 3000 Divisions = +/- 3 Divisions
Applied Load	As Found (kg)	Definitive (kg)	
10,000 kg	10,020 kg	10,020 kg	
10,000 kg	10,020kg	10,020 kg	
10,000 kg	10,020kg	10,020 kg	

Engineer:	C. Koning
Calibration Date:	26/09/2025
Calibration Renewal Date:	Sept 2026



All Calibrations are performed to UKWF Guidelines